

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

Packaged Product:				
<b>-</b> .	SODIUM N	METHOXIDE 0.1 NORMAL I		<b>.</b>
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	12/17/24			
Tester	ENOR			
Recorded	12/17/24			
NORMALITY TO 4 DEC I	PLACES	0.1000	0.0992 - 0.1008	0.1000
		EQUIV/LITER		
Sample ID				
Tested	12/17/24			
Tester	ENOR			
Recorded	12/17/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
	-	APPEARANCE		
Sample ID		/		
Tested	12/17/24			
Tester	ENOR			
Recorded	12/17/24			
Actual Density	12/17/24	0.79677	0.79610 - 0.79744	0.79702
Actual Density		G/ML	0.73010 0.73744	0.10102
O a marke ID		G/WE		
Sample ID				
Tested	12/17/24			
Tester	ENOR			
Recorded	12/17/24			
INSTRUMENTS USED D	URING	INSTRUMENT		M112,VF5-L14,M55
PREPARATION				
Sample ID				
Tested	12/17/24			
Tester	ENOR			
Recorded	12/17/24			
EXPIRATION DATE		MM/DD/YY		06/30/2025
		MM/DD/YYYY		
Sample ID				
Tested	12/17/24			
Tester	ENOR			
Recorded	12/17/24			
Lot # 41204	02			
Made 12/17	/ 2 4			



Page: 2 Date: 12/17/24 at 1:46 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

11hard SE

**Quality Manager**