



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ND720			
Heptane, Laboratory Grade 0.2 uM Filtered (4LG)			
Lot # 4120612			
Chemical Product: H1905			
HEPTANE LAB GRADE			
Test	Target/UOM	Range	Result
SPECIFIC ROTATION [A] AT 25 DEG	0.7223		0.70
	DEGREES		
Sample ID			
Tested 08/06/24			
Tester TBUR			
Recorded 08/06/24			
COLOR	> 28		30
	PT-CO		
Sample ID			
Tested 08/06/24			
Tester TBUR			
Recorded 08/06/24			
ANILINE POINT	PASS		PASS
	ANILINE		
Sample ID			
Tested 08/06/24			
Tester TBUR			
Recorded 08/06/24			
DISTILLATION TEMP IBP °C	>= 92		94
	TEMP		
Sample ID			
Tested 08/06/24			
Tester TBUR			
Recorded 08/06/24			
DISTILLATION TEMP DP °C	<=100		99
	TEMP		
Sample ID			
Tested 08/06/24			
Tester TBUR			
Recorded 08/06/24			
APPEARANCE AND COLOR	PASSES TEST		PASSES TEST
	APPEARANCE		
Sample ID			
Tested 08/06/24			
Tester TBUR			
Recorded 08/06/24			
Made 08/01/24			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 12/18/24 at 4:35 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager