

Made 12/19/24

Page: 1

Date: 12/19/24 at 4:26 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged				
_	TISAB	SOLUTION APHA-EPA FOR FLU		
Test		Target/UOM	Range	Result
Actual pH of	product	5.300	5.000 - 5.500	5.260
		PH		
	Sample ID			
	Tested 12/19/24			
	Tester JENC			
	Recorded 12/19/24			
PREPARED	TO FORMULATION ON	YES		YES
ILE		YES		
	Sample ID			
	Tested 12/19/24			
	Tester JENC			
	Recorded 12/19/24			
APPEARANC	CE AND COLOR	CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
	Sample ID	, <u>_</u> ,		
	Tested 12/19/24			
	Tester JENC			
	Recorded 12/19/24			
	ITS USED DURING	INSTRUMENT		VF12-100,GC,M113,T-26
PREPARATION		INSTITUMENT		V1 12-100,GO,W113,1-20
	Sample ID			
	Tested 12/19/24			
	Tester JENC			
Recorded 12/19/24				10/00/07
EXPIRATION	IDATE	MM/DD/YY		12/30/25
		MM/DD/YYYY		
	Sample ID			
	Tested 12/19/24			
	Tester JENC			
	Recorded 12/19/24			
SHELF LIFE	OF PRODUCT	SHELF LIFE		1 YEAR
		TIME		
	Sample ID			
	Tested 12/19/24			
	Tester JENC			
	Recorded 12/19/24			
LOURIDE L		0.2	0.0 - 0.5	0.0
		PPM		
	Sample ID			
	Tested 12/19/24			
	Tester JENC			
	Recorded 12/19/24			
	110001000 12/13/24			



Page: 2

Date: 12/19/24 at 4:26 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager