

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	1795			
	CHLORIN	NE STANDARD 1000 PPM AS	CL2 (Refrigerate)	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES	~	YES
FILE		YES		
Sample ID				
Tested	12/26/24			
Tester	ENOR			
Recorded	12/26/24			
PPM CHLORINE (CL2)		1000.0	995.0 - 1005.0	1004.3
		PPM		
Sample ID				
Tested	12/26/24			
Tester	ENOR			
Recorded	12/26/24			
APPEARANCE AND CO	LOR	CLEAR PALE YELLOW		PASS
		APPEARANCE		
Sample ID				
Tested	12/26/24			
Tester	ENOR			
Recorded	12/26/24			
Actual Density		1.00100	0.99969 - 1.00230	1.00120
		G/ML		
Sample ID				
Tested	12/26/24			
Tester	ENOR			
Recorded	12/26/24			
INSTRUMENTS USED D		INSTRUMENT		M112,VF4-N10,BID-08446,PID-
PREPARATION				361
Sample ID				001
Tested	12/26/24			
Tester	ENOR			
Recorded	12/26/24			
EXPIRATION DATE	12/20/24	MM/DD/YY		03/03/25
		MM/DD/YYYY		00/00/20
Sample ID				
Tested	10/06/04			
_	12/26/24			
Tester Recorded	ENOR 12/26/24			
SHELF LIFE OF PRODU		2 MONTHS		2 MONTHS
		TIME		
Sample ID				
Tested	10/06/04			
Tester	12/26/24 ENOR			
Recorded	ENOR 12/26/24			
Lot # 41220				
Made 12/26	5/24			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

11hard SE

Quality Manager