



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8800			
SODIUM THIOSULFATE 0.025 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.0250 EQ/L	0.0245 - 0.0255	0.0251
Sample ID			
Tested	12/20/24		
Tester	LWIL		
Recorded	12/20/24		
QCSP10:STANDARDIZATION OF SODIUM THIOSUL	YES QCSP10		YES
Sample ID			
Tested	12/20/24		
Tester	LWIL		
Recorded	12/20/24		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	12/20/24		
Tester	LWIL		
Recorded	12/20/24		
Actual Density	1.00163 G/ML	1.00140 - 1.00185	1.00183
Sample ID			
Tested	12/20/24		
Tester	LWIL		
Recorded	12/20/24		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	12/20/24		
Tester	LWIL		
Recorded	12/20/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,M91,T100A,M55,P106
Sample ID			
Tested	12/20/24		
Tester	LWIL		
Recorded	12/20/24		
TRACEABLE TO NIST SRM 136	YES YES		YES
Sample ID			
Tested	12/20/24		
Tester	LWIL		
Recorded	12/20/24		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		06/30/25
Sample ID			
Tested	12/20/24		
Tester	LWIL		
Recorded	12/20/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 12/20/24 at 1:56 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

6 MONTHS
TIME

6 MONTHS

Sample ID

Tested 12/20/24

Tester LWIL

Recorded 12/20/24

Lot # 4121027

Made 12/20/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager