



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: M3111			
METHYL tert-BUTYL ETHER REAGENT ACS GRADE			
Test	Target/UOM	Range	Result
ASSAY OF MATERIAL	MIN 99%		>99.99%
	%		
Sample ID			
Tested	12/17/24		
Tester	TBUR		
Recorded	12/17/24		
COLOR	MAX 10 APHA PT-CO		<10 APHA
Sample ID			
Tested	12/17/24		
Tester	TBUR		
Recorded	12/17/24		
PEROXIDE	MAX 1 PPM		<0.1 PPM
	%		
Sample ID			
Tested	12/17/24		
Tester	TBUR		
Recorded	12/17/24		
WATER CONTENT	MAX 0.05%		0.010%
	%		
Sample ID			
Tested	12/17/24		
Tester	TBUR		
Recorded	12/17/24		
RESIDUE AFTER EVAPORATION	MAX 0.001%		<0.0001%
Sample ID			
Tested	12/17/24		
Tester	TBUR		
Recorded	12/17/24		
Lot #	434411		
Made	12/09/24		

Richard [Signature]

Quality Manager