

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	• BH001F			
i uokugea i rouuoti	•	nine Stock Standard 1	.000 ppm w/v Refrigerate (1	Τ.)
Test	neenyian			
Test		Target/UOM YES	Range	Result YES
PREPARED TO FORMU	JLATION ON			YES
FILE		YES		
Sample ID				
Tested	12/18/24			
Tester	MMAT			
Recorded	12/18/24			1000 /
CALCULATED PPM ME	THYLAMINE	1000.0	996.0 - 1004.0	1000.1
		PPM		
Sample ID				
Tested	12/18/24			
Tester	MMAT			
Recorded	12/18/24			
APPEARANCE AND CO	LOR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	12/18/24			
Tester	MMAT			
Recorded	12/18/24			
INSTRUMENTS USED DURING		INSTRUMENT		VF1-Z6
PREPARATION				
Sample ID				
Tested	12/18/24			
Tester	MMAT			
Recorded	12/18/24			
EXPIRATION DATE		MM/DD/YYYY		02/23/2025
		MM/DD/YYYY		
Sample ID				
Tested	12/18/24			
Tester	MMAT			
Recorded	12/18/24			
Lot # 41209138				
Made 12/18	3/24			
inado				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA