

Page: 1

Date: 12/31/24 at 1:46 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	209919F			
	Titratio	on Solution with SN 0	.282 N AF Grade	
Test		Target/UOM	Range	Result
PREPARED TO FORMULA	ATION ON	YES	y	YES
FILE		YES		
Sample ID				
Tested	12/31/24			
Tester	RIRO			
Recorded	12/31/24			
NORMALITY TO 4 DEC PL	LACES	0.2820	0.2815 - 0.2825	0.2817
		EQUIV/LITER		
Sample ID				
· ·	12/31/24			
	RIRO			
Recorded	12/31/24			
QCSP1:STADARDIZATION	N OF	YES		YES
SILVER NITRATE		QCSP1		
Sample ID				
·	12/31/24			
	RIRO			
	12/31/24			
APPEARANCE AND COLO	OR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
·	12/31/24			
	RIRO			
	12/31/24			
INSTRUMENTS USED DU	RING	INSTRUMENT		M103,VF12-5,GC,BID-00488,PI
PREPARATION				D-M91
Sample ID				
	12/31/24			
	RIRO			
	12/31/24			
Actual Density		0.99596	0.99200 - 0.99991	0.99377
,		G/ML		
Sample ID				
·	12/31/24			
	RIRO			
	12/31/24			
EXPIRATION DATE		MM/DD/YY		12/30/26
		MM/DD/YYYY		
Sample ID				
	12/31/24			
	RIRO			
	12/31/24			
Lot # 4121343				
Made 12/31/	۷ ⁴			



Page: 2

Date: 12/31/24 at 1:46 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager