

Page: 1

Date: 01/03/25 at 8:00 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	1760			
	CERIC	SULFATE TITRANT 0.1 NORM	AL	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	ILATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	01/03/25			
Tester	LWIL			
Recorded	01/03/25			
Actual Normality		0.1000	0.0995 - 0.1005	0.1001
•		EQ/L		
Sample ID				
Tested	01/03/25			
Tester	LWIL			
Recorded	01/03/25			
Actual Density	01/00/20	1.08922	1.08601 - 1.09243	1.08963
riotaar Bonorty		G/ML	1.00001 1.00210	1.00000
Sample ID		S/ME		
Tested	01/03/25			
Tester	LWIL			
Recorded	01/03/25			
		YES		YES
QCSP19:STANDARDIZATION OF CERIC SULFATE		QCSP19		123
		QCSF 19		
Sample ID	0.1.10.0.10.7			
Tested	01/03/25			
Tester	LWIL			
Recorded	01/03/25	0.545.05440544045		B. 60
APPEARANCE AND CO	LOR	CLEAR ORANGE LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	01/03/25			
Tester	LWIL			
Recorded	01/03/25			
INSTRUMENTS USED I	DURING	INSTRUMENT		M113,T-230,VF2-A8,M55
PREPARATION				
Sample ID				
Tested	01/03/25			
Tester	LWIL			
Recorded	01/03/25			
EXPIRATION DATE		MM/DD/YY		07/30/25
		MM/DD/YYYY		
Sample ID				
Tested	01/03/25			
Tester	LWIL			
Recorded	01/03/25			
Lot # 41211				
Made 01/03	3/25			



Page: 2

Date: 01/03/25 at 8:00 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager