

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Package	ed Product:	0305			
0		ALKALINI	TY STANDARD 100 PPM		
Test			Target/UOM	Range	Result
	D TO FORMULAT	TION ON	YES		YES
FILE					
	Sample ID				
	Tested 12	2/31/24			
	Tester JF	NO			
	Recorded 12	2/31/24			
Alkalinity			100.00	90.00 - 110.00	94.89
			ALKALINITY		
	Sample ID				
1	Tested 12	2/31/24			
	Tester JF	RIO			
		2/31/24			
IC TO 2 DECIMAL PLACES			12.00	9.00 - 15.00	12.15
			PPM		
	Sample ID				
	Tested 12	2/31/24			
	Tester JF	OIF			
		2/31/24			
APPEARANCE AND COLOR		R	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	-	2/31/24			
		OIF			
		2/31/24			
Actual Der	nsity		0.99832	0.99822 - 0.99842	0.99833
			G/ML		
	Sample ID				
1		2/31/24			
		RIO			
		2/31/24			
INSTRUM	ENTS USED DUR	ING	INSTRUMENT		M91,VF12-101,M-66,BURET:07
PREPARA					9
	Sample ID				
	-	2/31/24			
		RIO			
		2/31/24			
SHELF LIF			3 MONTHS		3 MONTHS
			TIME		
	Sample ID				
	-	2/31/24			
		RIO			
		2/31/24			
EXPIRATIO			MM/DD/YYYY		03/30/2025
			MM/DD/YYYY		
	Sample ID				
1		2/31/24			
		RIO			
I		2/31/24			
I.					



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4121394
Made	12/31/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

chard St

Quality Manager