

Page: 1

Date: 01/04/25 at 6:54 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	1734			
	CERIC	SULFATE 0.01 NORMAL SOLU	TTE 0.01 NORMAL SOLUTION	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	01/04/25			
Tester	ENOR			
Recorded	01/04/25			
NORMALITY TO 4 DEC PLACES		0.0100	0.0097 - 0.0103	0.0102
		EQUIV/LITER		
Sample ID				
Tested	01/04/25			
Tester	ENOR			
Recorded	01/04/25			
QCSP19:STANDARDIZATION OF		YES		YES
CERIC SULFATE		QCSP19		
Sample ID				
Tested	01/04/25			
Tester	ENOR			
Recorded	01/04/25			
APPEARANCE AND CO	DLOR	CLEAR ORANGE LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	01/04/25			
Tester	ENOR			
Recorded	01/04/25			
Actual Density		1.16309	1.15052 - 1.17566	1.16997
		G/ML		
Sample ID				
Tested	01/04/25			
Tester	ENOR			
Recorded	01/04/25			
INSTRUMENTS USED	DURING	INSTRUMENT		M91,VF12-102
PREPARATION				
Sample ID				
Tested	01/04/25			
Tester	ENOR			
Recorded	01/04/25			
EXPIRATION DATE		MM/DD/YY		07/30/25
		MM/DD/YYYY		
Sample ID				
Tested	01/04/25			
Tester	ENOR			
Recorded	01/04/25			
Lot # 4121				
Made 01/0	4/20			

Page: 2

Date: 01/04/25 at 6:54 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager