Page: 1

Date: 01/07/25 at 10:08 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: GFS-98981			
_	KF WATER STANDA	ARD 0.05mg/g (50ppm) (2311)(10x5ML)	
Test	Targ	et/UOM	Range	Result
WATER CONTENT	RE	PORT		0.051
		%		
Sample ID				
Tested	01/07/25			
Tester	TBUR			
Recorded	01/07/25			
TRACEABLE TO N.I.S.		PORT		SRM 2890
		YES		
Sample ID				
Tested	01/07/25			
Tester	TBUR			
Recorded	01/07/25			
EXPIRATION DATE		DD/YYYY		06/19/2029
	MM/I	DD/YYYY		
Sample ID				
Tested	01/07/25			
Tester	TBUR			
Recorded	01/07/25			
Lot # 435105				
Made 12/16/24				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager