

Page: 1

Date: 01/14/25 at 1:46 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

			Her 1 0 # .	
Packaged Product:	CY157F			
	CYANIDE	STANDARD 100 PPM		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES	- 3	YES
FILE		YES		
Sample ID				
Tested	01/14/25			
Tester	RIRO			
Recorded	01/14/25			
Actual PPM		100.00	98.00 - 102.00	99.25
		PPM		
Sample ID				
Tested	01/14/25			
Tester	RIRO			
Recorded	01/14/25			
QCSP47:STANDARDIZATION		YES		YES
VARIOUS CYANIDE		QCSP47		
Sample ID				
Tested	01/14/25			
Tester	RIRO			
Recorded	01/14/25			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	01/14/25			
Tester	RIRO			
Recorded	01/14/25			
INSTRUMENTS USED D	URING	INSTRUMENT		VF100-K12,BID00488,TID4112
PREPARATION				4
Sample ID				
Tested	01/14/25			
Tester	RIRO			
Recorded	01/14/25			
SHELF LIFE OF PRODUCT		6 WEEKS		6 WEEKS
		TIME		
Sample ID				
Tested	01/14/25			
Tester	RIRO			
Recorded	01/14/25			
EXPIRATION DATE		MM/DD/YYYY		02/25/2025
		MM/DD/YYYY		
Sample ID				
Tested	01/14/25			
Tester	RIRO			
Recorded	01/14/25			
<b>Lot #</b> 50113				
<b>Made</b> 01/14	/25			



Page: 2

Date: 01/14/25 at 1:46 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**