Page: 1

Date: 01/07/25 at 4:09 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packag	ed Product:	CY369F			
		SKA 10,000	PPM NH3 STANDARD		
Test			Target/UOM	Range	Result
PREPARED TO FORMULATION ON			YES		YES
FILE			YES		
	Sample ID				
	Tested	01/07/25			
	Tester	MMAT			
	Recorded	01/07/25			
CALCULATED PPM NH3			10000	9975 - 10025	10001
			PPM		
	Sample ID				
		01/07/25			
		MMAT			
	Recorded	01/07/25			
APPEARANCE AND COLOR		OR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
		01/07/25			
		MMAT			
		01/07/25			
INSTRUMENTS USED DURING			INSTRUMENT		M120,DP,VF4-N32
PREPARATION					, ,
	Sample ID				
		01/07/25			
		MMAT			
		01/07/25			
EXPIRATION DATE		0.170.720	MM/DD/YYYY		04/30/2025
	.0		MM/DD/YYYY		0 1/100/2020
	Sample ID		14114, 25, 1111		
	•	01/07/25			
		MMAT			
		01/07/25			
L	ot # 501071	.02			
	ade 01/07/				
IVI	aue -/ -/				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I