

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Pr	oduct:	LY126F			
		CYANIDE S	TANDARD 5.0 PPM AQUEOUS		
Test			Target/UOM	Range	Result
PREPARED TO	FORMULA	TION ON	YES	-	YES
FILE			YES		
Sar	mple ID				
Tes	sted 01	1/13/25			
Tes	ster M	IMAT			
Re	corded 01	1/13/25			
APPEARANCE AND COLOR		R	CLEAR LIQUID		PASS
			APPEARANCE		
Sar	mple ID				
Tes	sted 01	1/13/25			
Tes	ster M	IMAT			
Re	corded 01	1/13/25			
INSTRUMENTS USED DURING			INSTRUMENT		P216,VF100-K4
PREPARATION					
Sar	mple ID				
Tes	sted 01	1/13/25			
Tes	ster M	IMAT			
Re	corded 01	1/13/25			
Actual ppm Cyanide			5.00	4.90 - 5.10	5.02
			PPM		
Sai	mple ID				
Tes	sted 01	1/13/25			
Tes	ster M	IMAT			
Re	corded 01	1/13/25			
EXPIRATION DATE			MM/DD/YY		03/18/25
			MM/DD/YYYY		
Sai	mple ID				
Tes	sted 01	1/13/25			
		IMAT			
Ree	corded 01	1/13/25			
Lot #	5011355	5			
	01/13/2				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA