

Recorded 01/17/25

Page: 1

Date: 01/17/25 at 3:01 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :					
Packaged Product: 9108					
SULFUR	IC ACID 7.0 NORMAL				
Test	Target/UOM	Range	Result		
PREPARED TO FORMULATION ON	YES		YES		
FILE	YES				
Sample ID					
Tested 01/17/25					
Tester JRIO					
Recorded 01/17/25					
SAMPLE COLLECTION & HANDLING	YES		YES		
PER LAB040	YES				
Sample ID					
Tested 01/17/25					
Tester JRIO					
Recorded 01/17/25					
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		01/17/2025		
PURPOSES	MM/DD/YYYY				
Sample ID					
Tested 01/17/25					
Tester JRIO					
Recorded 01/17/25					
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	20.0		
SAMPLING DEGC	DEG C				
Sample ID					
Tested 01/17/25					
Tester JRIO					
Recorded 01/17/25					
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		01/17/2025		
	MM/DD/YYYY				
Sample ID					
Tested 01/17/25					
Tester JRIO					
Recorded 01/17/25					
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.0		
DEG C	DEG C				
Sample ID					
Tested 01/17/25					
Tester JRIO					
Recorded 01/17/25					
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS		
	LIQUID				
	APPEARANCE				
Sample ID					
Tested 01/17/25					
Tester JRIO					
Recorded 01/17/25					
QCSP5: STANDARDIZATION OF	YES		YES		
AQUEOUS ACIDS	METHOD				
Sample ID					
Tested 01/17/25					
Tester JRIO					
Recorded 01/17/25					

Page: 2

Date: 01/17/25 at 3:01 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

INSTRUMENTS UPREPARATION	SED DURING	INSTRUMENT		VF3-J7,GC			
Sam	ple ID						
Teste	ed 01/17/25						
Teste	er JRIO						
Reco	orded 01/17/25						
NORMALITY TO 4	I DEC PLACES	7.0000	6.9900 - 7.0100	6.9907			
		EQUIV/LITER					
Samı	ple ID						
Teste	ed 01/17/25						
Teste	er JRIO						
Reco	orded 01/17/25						
Actual Density		1.20599	1.02504 - 1.20694	1.20675			
		G/ML					
Samı	ple ID						
Teste	ed 01/17/25						
Teste	er JRIO						
Reco	orded 01/17/25						
N.I.S.T. Traceable to SRM 723 THAM		YES		YES			
		YES					
Samı	ple ID						
Teste							
Teste	er JRIO						
Reco	orded 01/17/25						
EXPIRATION DAT	Έ	MM/DD/YY		01/30/27			
		MM/DD/YYYY					
Samı	ple ID						
Teste	ed 01/17/25						
Teste							
Reco	orded 01/17/25						
Lot # 5	0107104						
	Made 01/17/25						

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager