

**Made** 01/16/25

Page: 1

Date: 01/16/25 at 11:49 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

## Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	ct: 0373			
	AMMONIA	STANDARD 1,000 PPM AS	NH3 (AMMONIUM CHLORIDE)	
Test		Target/UOM	Range	Result
PREPARED TO FORM	JULATION ON	YES		YES
FILE		YES		
Sample I	D			
Tested	01/16/25			
Tester	MMAT			
Recorde	d 01/16/25			
SHELF LIFE OF PRODUCT		SHELF LIFE		3 MONTHS
		TIME		
Sample I	D			
Tested	01/16/25			
Tester	MMAT			
Recorde				
PPM NH3, CALCULA		1000	960 - 1040	1022
- <b>,</b> -		PPM		
Sample I	D			
Tested	01/16/25			
Tester	MMAT			
Recorde				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
AI I LANAITOL AITO	OLOR	APPEARANCE		17.00
Sample I	D	ALLEAGANOL		
Tested				
	01/16/25			
Tester	MMAT			
Recorde		INSTRUMENT		M409 CD M407
INSTRUMENTS USEI	DURING	INSTRUMENT		M108,GP,M107
PREPARATION	_			
Sample I				
Tested	01/16/25			
Tester	MMAT			
Recorde	d 01/16/25			
EXPIRATION DATE		MM/DD/YY		04/30/25
		MM/DD/YYYY		
Sample I	D			
Tested	01/16/25			
Tester	MMAT			
Recorde	d 01/16/25			
Lot # 501	 1514			
<b></b>	_			



Page: 2

Date: 01/16/25 at 11:49 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**