



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS121F Phosphate Standard 25.0 ppm Solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 01/10/25			
Tester MHUS			
Recorded 01/10/25			
PPM P04 TO 1 DECIMAL PLACE	25.0 PPM	24.5 - 25.5	25.0
Sample ID			
Tested 01/10/25			
Tester MHUS			
Recorded 01/10/25			
APPEARANCE AND COLOR	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 01/10/25			
Tester MHUS			
Recorded 01/10/25			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P117,VF1-Z39
Sample ID			
Tested 01/10/25			
Tester MHUS			
Recorded 01/10/25			
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		07/30/2025
Sample ID			
Tested 01/10/25			
Tester MHUS			
Recorded 01/10/25			
Lot #	5011056		
Made	01/10/25		

Richard [Signature]

Quality Manager