01/17/25

Made

Page: 1

Date: 01/17/25 at 4:14 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

		CaCO3 Sto	d. 625 mg/L For EZ4004-Total Alkalinity		
Test			Target/UOM	Range	Result
PREPARE	D TO FORMULA	TION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested 01	1/17/25			
	Tester J7	ΓRA			
	Recorded 01	1/17/25			
CALCULATED PPM AS CACO3		625.0	622.0 - 628.0	625.0	
(MG/L)			MG/L		
	Sample ID				
	Tested 01	1/17/25			
	Tester J1	ΓRA			
	Recorded 01	1/17/25			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS	
			APPEARANCE		
	Sample ID				
	Tested 01	1/17/25			
	Tester J1	ΓRA			
	Recorded 01	1/17/25			
INSTRUMENTS USED DURING		INSTRUMENT		M108,VF2-B13	
PREPARA	ATION				
	Sample ID				
		1/17/25			
	Tester J1	ΓRA			
	Recorded 01	1/17/25			
EXPIRATION DATE		MM/DD/YYYY		03/22/2025	
			MM/DD/YYYY		
	Sample ID				
		1/17/25			
	Tester J1	ΓRA			
		1/17/25			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I