

Tester

BPHI

Recorded 01/16/25

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	6918			
	CONDUC	TIVITY STANDARD 1.0 uMHOS		
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	01/16/25			
Tester	BPHI			
Recorded	01/16/25			
QCSP30:CONDUCTIVITY		YES		YES
STD&MEASUREMENT		QCSP30		
Sample ID				
	01/16/25			
	BPHI			
	01/16/25			
TRACEABLE TO N.I.S.T.		YES		YES
		YES		
Sample ID				
	01/16/25			
	BPHI			
	01/16/25			
APPEARANCE AND COLO		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
	04/40/05			
	01/16/25			
	BPHI			
	01/16/25			M00 T004
INSTRUMENTS USED DU	IRING	INSTRUMENT		M92,T234
PREPARATION				
Sample ID				
	01/16/25			
	BPHI			
	01/16/25			
SHELF LIFE		SHELFLIFE		3 MONTHS
		TIME		
Sample ID				
Tested	01/16/25			
Tester	BPHI			
Recorded	01/16/25			
EXPIRATION DATE		MM/DD/YY		04/16/25
		MM/DD/YYYY		
Sample ID				
	01/16/25			
	BPHI			
	01/16/25			
CONDUCTIVITY TO 1 DEC		1.0	0.9 - 1.1	1.0
	· · • •	US/CM		
Sample ID		00,011		
	01/16/05			
	01/16/25			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	5011648
Made	01/16/25

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

chard Sto

Quality Manager