Page: 1

Date: 01/08/25 at 5:19 PM

## **Aqua Solutions, Inc.** 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

	AMMONIA	STANDARD 1.0 PPM AS NH3		
Test		Target/UOM	Range	Result
PREPARED TO	FORMULATION ON	YES		YES
FILE		YES		
Sar	nple ID			
Tes	ted 01/08/25			
Tes	ter JTRA			
Red	orded 01/08/25			
CERTIFIED PPM NH3		1.0	0.8 - 1.2	1.1
		PPM		
Sar	nple ID			
Tes	•			
Tes				
Red	corded 01/08/25			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sar	nple ID			
Tes	ited 01/08/25			
Tes	ter JTRA			
Red	corded 01/08/25			
INSTRUMENTS USED DURING		INSTRUMENT		P390,M107,VF2-B18
PREPARATION				
Sar	nple ID			
Tes	ted 01/08/25			
Tes	ter JTRA			
Red	corded 01/08/25			
EXPIRATION DATE		MM/DD/YYYY		02/13/2025
		MM/DD/YYYY		
Sar	nple ID			
Tes	•			
Tes				
Red	orded 01/08/25			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland Is