

Recorded 01/14/25

Page: 1

Date: 01/14/25 at 11:26 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:							
Packaged Produc	t: 9040						
	SULFURIC	ACID 0.1 NORMAL NIST	TRACEABLE SOLUTION				
Test		Target/UOM	Range	Result			
DATE OF TESTING OF SAMPLE		MM/DD/YYYY		01/14/2025			
		MM/DD/YYYY					
Sample II	)						
Tested	01/14/25						
Tester	DTRA						
Recorded	01/14/25						
AMBIENT TEST TEMPERATURE		20.0	15.0 - 30.0	18.0			
DEG C		DEG C					
Sample II	)						
Tested	01/14/25						
Tester	DTRA						
Recorded	01/14/25						
Actual Normality		0.1000	0.0995 - 0.1005	0.1002			
		EQ/L					
Sample II	)						
Tested	01/14/25						
Tester	DTRA						
Recorded	01/14/25						
QCSP5: STANDARDIZ	ATION OF	YES		YES			
AQUEOUS ACIDS		METHOD					
Sample II	)						
Tested	01/14/25						
Tester	DTRA						
Recorded							
N.I.S.T. Traceable to S	SRM 84	Yes		YES			
		YES					
Sample II	)						
Tested	01/14/25						
Tester	DTRA						
Recorded							
PREPARED TO FORMULATION ON		YES		YES			
FILE		YES					
Sample II	)						
Tested	01/14/25						
Tester	DTRA						
Recorded							
SAMPLE COLLECTIO		YES		YES			
PER LAB040		YES					
Sample II	)						
Tested	01/14/25						
Tester	DTRA						
Recorded							
DATE OF SAMPLING FOR TEST		MM/DD/YYYY		01/14/2025			
PURPOSES		MM/DD/YYYY					
Sample II	)						
Tested	01/14/25						
Tester	DTRA						



Page: 2

Date: 01/14/25 at 11:26 AM

## **Aqua Solutions, Inc.** 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO #:

AMBIENT TEMPERATURE DURING		25.0	10.0 - 35.0	18.0
SAMPLING DEGC		DEG C		
Sample ID				
Tested	01/14/25			
Tester	DTRA			
Recorded	01/14/25			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	01/14/25			
Tester	DTRA			
Recorded	01/14/25			
INSTRUMENTS USED		INSTRUMENT		M55MGC,T203
PREPARATION				
Sample ID				
Tested	01/14/25			
Tester	01/14/25 DTRA			
Recorded	01/14/25	MANA/DD/XXX		04/20/27
EXPIRATION DATE		MM/DD/YY		01/30/27
		MM/DD/YYYY		
Sample ID				
Tested	01/14/25			
Tester	DTRA			
Recorded	01/14/25			
SHELF LIFE OF PROD	UCT	SHELF LIFE		2 YEARS
		TIME		
Sample ID				
Tested	01/14/25			
Tester	DTRA			
Recorded	01/14/25			
Actual Density		1.00169	1.00143 - 1.00194	1.00173
-		G/ML		
Sample ID				
Tested	01/14/25			
Tester	DTRA			
Recorded	01/14/25			
Lot # 5010				
Made 01/1	4/25			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Page: 3

Date: 01/14/25 at 11:26 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**