



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: HAC460F			
Calibration Soln 0.5 PPM CaCO3 for EZ1016-Total Hardness			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 01/13/25			
Tester MMAT			
Recorded 01/13/25			
CALCULATED PPM CaCO3	0.50	0.47 - 0.53	0.50
	PPM		
Sample ID			
Tested 01/13/25			
Tester MMAT			
Recorded 01/13/25			
APPEARANCE AND COLOR	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 01/13/25			
Tester MMAT			
Recorded 01/13/25			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P101,VF2-A8
Sample ID			
Tested 01/13/25			
Tester MMAT			
Recorded 01/13/25			
EXPIRATION DATE	MM/DD/YYYY		04/30/2025
	MM/DD/YYYY		
Sample ID			
Tested 01/13/25			
Tester MMAT			
Recorded 01/13/25			
Lot #	5011373		
Made	01/13/25		

Richard [Signature]
Quality Manager