



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8435			
SILVER NITRATE 0.05 NORMAL			
Test	Target/UOM	Range	Result
N.I.S.T. Traceable	Yes		YES
	YES		
Sample ID			
Tested 01/24/25			
Tester LWIL			
Recorded 01/24/25			
PREPARED TO FORMULATION ON FILE	YES		YES
	YES		
Sample ID			
Tested 01/24/25			
Tester LWIL			
Recorded 01/24/25			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
	YES		
Sample ID			
Tested 01/24/25			
Tester LWIL			
Recorded 01/24/25			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		01/24/2025
	MM/DD/YYYY		
Sample ID			
Tested 01/24/25			
Tester LWIL			
Recorded 01/24/25			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	18.1
	DEG C		
Sample ID			
Tested 01/24/25			
Tester LWIL			
Recorded 01/24/25			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		01/24/2025
	MM/DD/YYYY		
Sample ID			
Tested 01/24/25			
Tester LWIL			
Recorded 01/24/25			
AMBIENT TEST TEMPERATURE DEGC	20.0	15.0 - 30.0	18.1
	DEG C		
Sample ID			
Tested 01/24/25			
Tester LWIL			
Recorded 01/24/25			
QCSP53: SILVER NITRATE STD ON AUTO-TITRA	YES		YES
	METHOD		
Sample ID			
Tested 01/24/25			
Tester LWIL			
Recorded 01/24/25			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Actual Normality	0.0500	0.0495 - 0.0505	0.0501
	EQ/L		
Sample ID			
Tested	01/24/25		
Tester	LWIL		
Recorded	01/24/25		
Actual Density	1.00537	1.00522 - 1.00551	1.00533
	G/ML		
Sample ID			
Tested	01/24/25		
Tester	LWIL		
Recorded	01/24/25		
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested	01/24/25		
Tester	LWIL		
Recorded	01/24/25		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M114,T101A
Sample ID			
Tested	01/24/25		
Tester	LWIL		
Recorded	01/24/25		
EXPIRATION DATE	MM/DD/YY		07/30/26
	MM/DD/YYYY		
Sample ID			
Tested	01/24/25		
Tester	LWIL		
Recorded	01/24/25		
Lot #	5011369		
Made	01/24/25		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager