

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	CY229F				
	COD BLIN	D QA STANDARD	(ALL SAME LOT)		
Test		Target/UOM		Range	Result
PPM COD		82.0		19.0 - 145.0	36.3
		PPM			
Sample ID					
•	01/22/25				
Tester	DTRA				
Recorded	01/22/25				
APPEARANCE AND COL	.OR	CLEAR WATER WHIT	E		PASS
		LIQUID			
		APPEARANCE			
Sample ID					
•	01/22/25				
	DTRA				
Recorded	01/22/25				
INSTRUMENTS USED DU	URING	LAB EQUIPMENT			P47,VF1-Z39
PREPARATION		INSTRUMENT			
Sample ID					
-	01/22/25				
	DTRA				
Recorded	01/22/25				
EXPIRATION DATE		MM/DD/YYYY			03/23/2025
		MM/DD/YYYY			
Sample ID					
Tested	01/22/25				
Tester	DTRA				
Recorded	01/22/25				
Lot # 501203	16				
Made 01/22,					
	, 20				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1 chard Stof

Quality Manager