

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Produ	uct: 6950			
U		IVITY STANDARD 1,000 uMH	IOS	
Test		Target/UOM	Range	Result
PREPARED TO FOR	RMULATION ON	YES		YES
FILE		YES		
Sample	lD			
Tested	01/24/25			
Tester	BPHI			
Record	ed 01/24/25			
CONDUCTIVITY @25 C.		1000	999 - 1001	1000
		US/CM		
Sample				
Tested	01/24/25			
Tester	BPHI			
Record				
QCSP30:CONDUCTIVITY		YES		YES
STD&MEASUREME		QCSP30		
Sample	lD			
Tested	01/24/25			
Tester	BPHI			
Record				
TRACEABLE TO N.I	.S.T. SRM 3191	YES		YES
		YES		
Sample	: ID			
Tested	01/24/25			
Tester	BPHI			
Record				
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample	e ID			
Tested	01/24/25			
Tester	BPHI			
Record	ed 01/24/25			
INSTRUMENTS USED DURING		INSTRUMENT		M92,T229
PREPARATION				
Sample	lD			
Tested	01/24/25			
Tester	BPHI			
Record	ed 01/24/25			
EXPIRATION DATE		MM/DD/YY		01/30/26
		MM/DD/YYYY		
Sample	lD			
Tested	01/24/25			
Tester	BPHI			
Record				
SHELF LIFE OF PR	ODUCT	SHELF LIFE		1 YEAR
		TIME		
Sample	: ID			
Tested	01/24/25			
Tester	BPHI			
Record	ed 01/24/25			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	5012077
Made	01/24/25

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

thand St

Quality Manager