

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product				
	Oxalic .	Acid Solution for EZ103	35 Silica Analyzer	
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample ID	1			
Tested	01/24/25			
Tester	MMAT			
Recorded	01/24/25			
CALCULATED GPL OXALIC ACID		100.0	99.9 - 100.1	100.0
		GPL		
Sample ID)			
Tested	01/24/25			
Tester	MMAT			
Recorded	01/24/25			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID	1			
Tested	01/24/25			
Tester	MMAT			
Recorded	01/24/25			
INSTRUMENTS USED DURING		INSTRUMENT		M120,VF2-B17
PREPARATION				
Sample ID	1			
Tested	01/24/25			
Tester	MMAT			
Recorded	01/24/25			
EXPIRATION DATE		MM/DD/YYYY		01/30/2026
		MM/DD/YYYY		
Sample ID)			
Tested	01/24/25			
Tester	MMAT			
Recorded	01/24/25			
Lot # 5012	321			
Made 01/2				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA