



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: SH080F			
Anion 8Comp 1000ppm Acetate, Cl, SO4, NO3, Carbonate, NO2, Br, Form			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 01/23/25			
Tester MMAT			
Recorded 01/23/25			
APPEARANCE AND COLOR	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 01/23/25			
Tester MMAT			
Recorded 01/23/25			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M108,VF1-Z24
Sample ID			
Tested 01/23/25			
Tester MMAT			
Recorded 01/23/25			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		04/30/25
Sample ID			
Tested 01/23/25			
Tester MMAT			
Recorded 01/23/25			
PPM SODIUM ACETATE	1000.0 PPM	995.0 - 1005.0	1000.3
Sample ID			
Tested 01/23/25			
Tester MMAT			
Recorded 01/23/25			
PPM SODIUM CHLORIDE	1000.0 PPM	995.0 - 1005.0	999.9
Sample ID			
Tested 01/23/25			
Tester MMAT			
Recorded 01/23/25			
CALCULATED PPM SODIUM SULFATE	1000.0 PPM	995.0 - 1005.0	999.7
Sample ID			
Tested 01/23/25			
Tester MMAT			
Recorded 01/23/25			
PPM SODIUM NITRATE	1000.0 PPM	995.0 - 1005.0	1000.2
Sample ID			
Tested 01/23/25			
Tester MMAT			
Recorded 01/23/25			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

PPM SODIUM CARBONATE	1000.0	995.0 - 1005.0	999.7
	PPM		
Sample ID			
Tested	01/23/25		
Tester	MMAT		
Recorded	01/23/25		
PPM SODIUM NITRITE	1000.0	995.0 - 1005.0	1000.1
	PPM		
Sample ID			
Tested	01/23/25		
Tester	MMAT		
Recorded	01/23/25		
PPM SODIUM BROMIDE	1000.0	995.0 - 1005.0	999.8
	PPM		
Sample ID			
Tested	01/23/25		
Tester	MMAT		
Recorded	01/23/25		
PPM SODIUM FORMATE	1000.0	995.0 - 1005.0	1000.1
	PPM		
Sample ID			
Tested	01/23/25		
Tester	MMAT		
Recorded	01/23/25		
Lot #	5012189		
Made	01/23/25		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager