

Page: 1

Date: 01/31/25 at 12:59 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

## Certificate of Analysis / QC Results Customer PO #:

oustomer i o

Packaged Produc	t: 1763			
	CHLORIDE	E STANDARD 10.0 PPM W/W	AS CL (N.I.S.T)	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID	)			
Tested	01/31/25			
Tester	MMAT			
Recorded	01/31/25			
CHLORIDE PPM		10.0	9.0 - 11.0	9.6
		PPM		
Sample ID	)			
Tested	01/31/25			
Tester	MMAT			
Recorded				
APPEARANCE AND CO		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID	)			
Tested	01/31/25			
Tester	MMAT			
Recorded				
INSTRUMENTS USED		INSTRUMENT		P54,M107,VF12-100
PREPARATION	Dorinto	INCTITIONEIVI		1 04,141107, 41 12 100
Sample ID	<b>\</b>			
Tested	01/31/25			
Tester Recorded	MMAT 01/31/25			
	01/31/25	MM/DD/YY		07/30/25
EXPIRATION DATE				07/30/25
		MM/DD/YYYY		
Sample ID				
Tested	01/31/25			
Tester	MMAT			
Recorded		0115151155		0.1401/TU0
SHELF LIFE OF PROD	UCT	SHELF LIFE TIME		6 MONTHS
Sample ID	)			
Tested	01/31/25			
Tester	MMAT			
Recorded				
<b>Lot #</b> 5013				
Made 01/3				
IVIAUE 01/9	-, -0			



Page: 2

Date: 01/31/25 at 12:59 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**