

Page: 1

Date: 01/27/25 at 4:09 PM

PASS

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: 2161
COD STANDARD 100 PPM SOLUTION

Test Target/UOM Range Result
PREPARED TO FORMULATION ON YES YES
FILE YES

Sample ID Tested 01/27/25

Tested 01/27/25
Tester DTRA
Recorded 01/27/25

APPEARANCE AND COLOR CLEAR LIQUID
APPEARANCE

Sample ID
Tested 01/27/25

Tester DTRA Recorded 01/27/25

INSTRUMENTS USED DURING INSTRUMENT P101,VF1-Z40

PREPARATION

Sample ID Tested 01/2

Tested 01/27/25 Tester DTRA Recorded 01/27/25

SHELF LIFE OF PRODUCT 2 MONTHS 2 MONTHS

TIME

Sample ID

Tested 01/27/25
Tester DTRA
Recorded 01/27/25

EXPIRATION DATE MM/DD/YYYY 04/01/2025

MM/DD/YYYY

Sample ID

Tested 01/27/25
Tester DTRA
Recorded 01/27/25

PPM COD 100.0 97.0 - 103.0 99.1

PPM

Sample ID

Tested 01/27/25
Tester DTRA
Recorded 01/27/25

Lot # 5012732 **Made** 01/27/25



Page: 2

Date: 01/27/25 at 4:09 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager