

Page: 1

Date: 01/28/25 at 1:48 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	:: 0371			
	Ammonia	Nitrogen Standard 10 F	PPM as NH3	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	01/28/25			
Tester	MMAT			
Recorded	01/28/25			
CERTIFIED PPM NH3		10.0	9.0 - 11.0	10.8
		PPM		
Sample ID				
Tested	01/28/25			
Tester	MMAT			
Recorded	01/28/25			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	01/28/25			
Tester	MMAT			
Recorded	01/28/25			
INSTRUMENTS USED		INSTRUMENT		P176,GP,VF12-100,M107
PREPARATION				
Sample ID				
Tested	01/28/25			
Tester	MMAT			
Recorded	01/28/25			
EXPIRATION DATE	01/20/23	MM/DD/YY		04/30/25
EXPINATION DATE		MM/DD/YYYY		04/00/23
Sample ID		1VIIVI/DD/1111		
Sample ID Tested				
Tested	01/28/25 MMAT			
rester Recorded	01/28/25			
SHELF LIFE OF PROD		SHELF LIFE		3 MONTHS
SHELF LIFE OF PROD	001	TIME		3 WONTTIS
Sample ID				
Tested	01/28/25			
Tester	MMAT			
Recorded	01/28/25			
Lot # 5012883				
Made 01/28/25				



Page: 2

Date: 01/28/25 at 1:48 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager