Page: 1

Date: 01/31/25 at 11:55 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

		Standard 100.0 ppm (C Source Ethylene Glycol)		
Test	Target/UOM	Range	Result	
PREPARED TO FORMULATION			YES	
FILE	YES			
Sample ID				
Tested 01/3	1/25			
Tester DTR	A			
Recorded 01/3	1/25			
CERTIFIED PPM TOC	100.0	98.0 - 102.0	99.4	
	PPM			
Sample ID				
Tested 01/3	1/25			
Tester DTR	A			
Recorded 01/3	1/25			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS	
	APPEARANCE			
Sample ID				
Tested 01/3	1/25			
Tester DTR	A			
Recorded 01/3	1/25			
INSTRUMENTS USED DURING	G INSTRUMENT		P101,VF2-B16,M109	
PREPARATION				
Sample ID				
Tested 01/3	1/25			
Tester DTR				
Recorded 01/3	1/25			
EXPIRATION DATE	MM/DD/YYYY		04/30/2025	
	MM/DD/YYYY			
Sample ID				
Tested 01/3	1/25			
Tester DTR				
Recorded 01/3				
Lot # 5012954				
Made 01/31/25				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I