

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product: 8360				
achagoa i readoti	R NITRATE 0.1 NORMAL IN	NITRATE 0.1 NORMAL IN IPA (ASTM D3227-13)		
Test	Target/UOM	Range	Result	
Actual Normality	0.1000	0.0995 - 0.1005	0.1001	
	EQ/L			
Sample ID				
Tested 01/30/25				
Tester RIRO				
Recorded 01/30/25				
PREPARED TO FORMULATION ON	YES		YES	
FILE	YES			
Sample ID				
Tested 01/30/25				
Tester RIRO				
Recorded 01/30/25				
SAMPLE COLLECTION & HANDLING	G YES		YES	
PER LAB040	YES			
Sample ID				
Tested 01/30/25				
Tester RIRO				
Recorded 01/30/25				
AMBIENT TEMPERATURE DURING	20.0	15.0 - 30.0	20.0	
SAMPLING DEGC	DEG C			
Sample ID				
Tested 01/30/25				
Tester RIRO				
Recorded 01/30/25				
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		01/30/2025	
	MM/DD/YYYY			
Sample ID				
Tested 01/30/25				
Tester RIRO				
Recorded 01/30/25				
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.0	
DEG C	DEG C			
Sample ID				
Tested 01/30/25				
Tester RIRO				
Recorded 01/30/25				
QCSP53: SILVER NITRATE STD ON	YES		YES	
AUTO-TITRA	METHOD			
Sample ID				
Tested 01/30/25				
Tester RIRO				
Recorded 01/30/25				
N.I.S.T TRACEABLE TO SRM 919	YES		YES	
	YES		120	
Sample ID	120			
-				
Tester RIRO				
Recorded 01/30/25				



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

		•••••		
IPA TESTED FOR PERC	DXIDES	PASSES		PASS
		YES		
Sample ID				
Tested	01/30/25			
Tester	RIRO			
Recorded	01/30/25			
Actual Density		0.83208	0.83068 - 0.83349	0.83132
		G/ML		
Sample ID				
Tested	01/30/25			
Tester	RIRO			
Recorded	01/30/25			
APPEARANCE AND CO		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	01/20/25			
	01/30/25			
Tester	RIRO 01/30/25			
INSTRUMENTS USED D	DURING	INSTRUMENT		T-A7,M112,M91,BURETM52,TI
PREPARATION				41124
Sample ID				
Tested	01/30/25			
Tester	RIRO			
Recorded	01/30/25			
EXPIRATION DATE				07/30/2026
		MM/DD/YYYY		
Sample ID				
Tested	01/30/25			
Tester	RIRO			
Recorded	01/30/25			
SHELF LIFE OF PRODU	JCT	SHELF LIFE		18 MONTHS
		TIME		
Sample ID				
Tested	01/30/25			
Tester	RIRO			
Recorded	01/30/25			
DATE OF SAMPLING FO	OR TEST	MM/DD/YYYY		01/30/2025
PURPOSES		MM/DD/YYYY		
Sample ID				
Tested	01/30/25			
Tester	RIRO			
Recorded	01/30/25			
Lot # 50115	548			
Made 01/30				
	125			



Page: 3 Date: 01/30/25 at 5:37 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

11hard SE

Quality Manager