

Page: 1

Date: 01/30/25 at 7:59 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product				
Took	TOTAL CH	HLORINE STANDARD 175 P		Danul
Test		Target/UOM	Range	Result YES
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	01/30/25			
Tester	BPHI			
Recorded	01/30/25			
PPM CHLORINE (CL2)		175.00	170.00 - 180.00	179.70
		PPM		
Sample ID				
Tested	01/30/25			
Tester	BPHI			
Recorded	01/30/25			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	01/30/25			
Tester	BPHI			
Recorded	01/30/25			
INSTRUMENTS USED I		INSTRUMENT		M91,VF1-240,M55,P54,BURET
PREPARATION				08446
Sample ID				00440
Tested	01/30/25			
Tester	01/30/25 BPHI			
Recorded	01/30/25			
	01/30/25	0.99851	0.99810 - 0.99891	0.99848
Actual Density		0.99651 G/ML	0.99610 - 0.99691	0.99646
0		G/IVIL		
Sample ID				
Tested	01/30/25			
Tester	BPHI			
Recorded	01/30/25			
EXPIRATION DATE		MM/DD/YYYY MM/DD/YYYY		03/06/2025
Sample ID				
Tested	01/30/25			
Tester	BPHI			
Recorded	01/30/25			
Lot # 5013	072			
Made 01/3				



Page: 2

Date: 01/30/25 at 7:59 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**