



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

502805
HA-22540-69

Certificate of Analysis

PRODUCT: Amino Acid F Reagent

PRODUCT NUMBER: 2254069

LOT NUMBER: A4363

MANUFACTURE DATE: 01/17/2025

DATE OF ANALYSIS: 01/17/2025

TEST	SPECIFICATIONS	RESULTS
Absorbance of a 1 ppm Silica standard vs the reagent blank.	0.336 to 0.394 abs	0.3890 abs
Absorbance of a reagent blank vs DI water at 700 nm in 1 inch cells.	-0.012 to 0.012 abs	0.0110 abs
Absorbance of a 3 ppm Silica standard vs the reagent blank.	1.022 to 1.168 abs	1.1030 abs

The expiration date is Dec 2027

Certified by: *Scott Als*

Analytical Services Chemist