



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

502003
HA-26605-49

Certificate of Analysis

PRODUCT: StablCal® Standard, 800 NTU

PRODUCT NUMBER: 2660549

LOT NUMBER: A4313

MANUFACTURE DATE: 11/21/2024

DATE OF ANALYSIS: 11/25/2024

TEST	SPECIFICATIONS	RESULTS
Turbidity	760 to 840 NTU	803.0 NTU

The expiration date is Nov 2026

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by: *Scott Als*

Analytical Services Chemist