



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500613  
HA-20760-30

### Certificate of Analysis

**PRODUCT:** Molybdovanadate Reagent

**PRODUCT NUMBER:** 2076032

**LOT NUMBER:** A4218

**MANUFACTURE DATE:** 08/15/2024

**DATE OF ANALYSIS:** 08/16/2024

TEST	SPECIFICATIONS	RESULTS
Function Test: Absorbance of phosphate standards with concentrations of 50.0 ppm PO4 vs the reagent blank.	0.483 to 0.549 abs	0.5199 abs
Function Test: Absorbance of phosphate standards with concentrations of 10.0 ppm PO4 vs the reagent blank.	0.1 to 0.122 abs	0.1059 abs
Reagent Blank: Measured at 500 nm in 5 cm vs DI water.	0 to 0.037 abs	0.0024 abs

The expiration date is Aug 2029

Certified by: *Scott Als*

Analytical Services Chemist