



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500623  
HA-23864-42

### Certificate of Analysis

**PRODUCT:** Amino Acid F Reagent Solution

**PRODUCT NUMBER:** 2386442

**LOT NUMBER:** Λ4309

**MANUFACTURE DATE:** 11/19/2024

**DATE OF ANALYSIS:** 11/19/2024

TEST	SPECIFICATIONS	RESULTS
pH of the solution	7.1 to 7.7	7.38
Absorbance of Silica standard vs the reagent blank with 2.0 ppm SiO <sub>2</sub> .	0.692 to 0.718 abs	0.7070 abs
Reagent Blank: Measured at 815 nm in 1 cm cell vs DI water.	0 to 0.005 abs	0.0010 abs

The expiration date is Jul 2025

Certified by: *Scott Als*

Analytical Services Chemist