

Honeywell Specialty Chemicals Seelze GmbH D-30918 Seelze
Telefon: +49 5137 999-353

501302
S0630

13.03.2024 / COAH_01098445

Analyzed for:
Honeywell PMT Advanced Materials E&C
Offering Management
115 Tabor Road, Morris Plains
NJ 07950, USA

QC Release Date: 13.Mar.2024
Production date: 14.Mar.2022
Expiry Date: 19.Sep.2026

Article/Product: 37865
HYDRANAL-Salicylic acid

Lot: M073C

The quality certificate is valid for the time of delivery

HYDRANAL Salicylic Acid is a special salicylic acid quality for Karl Fischer titration.

Parameter	Specification	QC results
Suitability (for KF-Titration)*	test	test passed
Assay of water*	maximum 0,2 %	0.022 %
Number of samples (n)	minimum 2	2

* volumetric KF titration according to Test Method 01 HYDRANAL version 13. Nov. 2017 (including ISO 760 version 1978 and ASTM E 203 version 2016).

The KF suitability test was performed on the first bottle from the filling line.

The assay of water is the highest value from a series of measurements on (n) samples.

- The expiry date is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.
- We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.
- Particular properties of the products or the suitability for a particular area of application are not assured.
- We guarantee a proper quality within our General Conditions of Sales.

Thomas Wendt

Honeywell Specialty Chemicals Seelze GmbH
Thomas Wendt
Supervisor HYDRANAL Center of Excellence
Wunstorfer Str. 40, D-30926 Seelze



Quality Testing Lab Accredited According to DIN EN ISO/IEC 17025

