



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

502016  
HA-21073-69

### Certificate of Analysis

**PRODUCT:** Molybdate Reagent

**PRODUCT NUMBER:** 2107369

**LOT NUMBER:** A4318

**MANUFACTURE DATE:** 11/22/2024

**DATE OF ANALYSIS:** 11/26/2024

TEST	SPECIFICATIONS	RESULTS
Function Test: Absorbance of 50 ppm silica standard vs the reagent blank measured at 410 nm in 2 mm cells.	0.248 to 0.296 abs	0.2590 abs
Reagent Blank: Absorbance measured at 410 nm in 1 cm cells vs deionized water.	0 to 0.011 abs	0.0020 abs

The expiration date is Nov 2029

Certified by: *Scott Als*

Analytical Services Chemist