



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8810			
SODIUM THIOSULFATE 0.1 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 02/07/25			
Tester JRIO			
Recorded 02/07/25			
PH OF PRODUCT/SOLUTION	7.000	6.000 - 12.500	10.098
	PH		
Sample ID			
Tested 02/07/25			
Tester JRIO			
Recorded 02/07/25			
Actual Normality	0.1000	0.0995 - 0.1005	0.1000
	EQ/L		
Sample ID			
Tested 02/07/25			
Tester JRIO			
Recorded 02/07/25			
QCSP10:STANDARDIZATION OF SODIUM THIOSUL	YES		YES
	QCSP10		
Sample ID			
Tested 02/07/25			
Tester JRIO			
Recorded 02/07/25			
QCSP35:STANDARDIZATION OF SODIUM THIOSUL	YES		YES
	QCSP35		
Sample ID			
Tested 02/07/25			
Tester JRIO			
Recorded 02/07/25			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 02/07/25			
Tester JRIO			
Recorded 02/07/25			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M113,M103,T-211,M106,BID07
			9
Sample ID			
Tested 02/07/25			
Tester JRIO			
Recorded 02/07/25			
Actual Density	1.01123	1.01079 - 1.01166	1.01121
	G/ML		
Sample ID			
Tested 02/07/25			
Tester JRIO			
Recorded 02/07/25			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

DATE OF PREPARATION	MM/DD/YYYY	02/07/2025
Sample ID		
Tested	02/07/25	
Tester	JRIO	
Recorded	02/07/25	
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	08/30/25
Sample ID		
Tested	02/07/25	
Tester	JRIO	
Recorded	02/07/25	
SHELF LIFE	6 MONTHS TIME	6 MONTHS
Sample ID		
Tested	02/07/25	
Tester	JRIO	
Recorded	02/07/25	
TRACEABLE TO NIST SRM 136	YES YES	YES
Sample ID		
Tested	02/07/25	
Tester	JRIO	
Recorded	02/07/25	
Lot #	5013051	
Made	02/07/25	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager