



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8410			
SILVER NITRATE 0.0282 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 02/07/25			
Tester JENC			
Recorded 02/07/25			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 02/07/25			
Tester JENC			
Recorded 02/07/25			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		02/07/2025
Sample ID			
Tested 02/07/25			
Tester JENC			
Recorded 02/07/25			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	19.8
Sample ID			
Tested 02/07/25			
Tester JENC			
Recorded 02/07/25			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		02/07/2025
Sample ID			
Tested 02/07/25			
Tester JENC			
Recorded 02/07/25			
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	19.8
Sample ID			
Tested 02/07/25			
Tester JENC			
Recorded 02/07/25			
Actual Normality	0.0282 EQ/L	0.0279 - 0.0285	0.0283
Sample ID			
Tested 02/07/25			
Tester JENC			
Recorded 02/07/25			
QCSP54:STANDARDIZATION OF LOW NORMAL SN	YES QCSP54		YES
Sample ID			
Tested 02/07/25			
Tester JENC			
Recorded 02/07/25			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 919 NaCl	YES YES	YES
Sample ID		
Tested	02/07/25	
Tester	JENC	
Recorded	02/07/25	
Actual Density	1.00225 G/ML	1.00205 - 1.00245 1.00220
Sample ID		
Tested	02/07/25	
Tester	JENC	
Recorded	02/07/25	
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE	PASS
Sample ID		
Tested	02/07/25	
Tester	JENC	
Recorded	02/07/25	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M112,T-260,M55,BURET:0641 9,M91
Sample ID		
Tested	02/07/25	
Tester	JENC	
Recorded	02/07/25	
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY	02/28/2027
Sample ID		
Tested	02/07/25	
Tester	JENC	
Recorded	02/07/25	
Lot #	5013127	
Made	02/07/25	

Richard [Signature]

Quality Manager