

Page: 1

Date: 02/04/25 at 11:32 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

		Ousid	mer row.	
Packaged Produ				
	CYANIDE	STANDARD 100.0 PPM (N		
Test		Target/UOM	Range	Result
PREPARED TO FOR	MULATION ON	YES		YES
FILE		YES		
Sample				
Tested	02/04/25			
Tester	BKIR			
Recorde				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample				
Tested	02/04/25			
Tester	BKIR			
Recorde	d 02/04/25	100.00	09.00 100.00	100.00
Actual ppm Cyanide		100.00 PPM	98.00 - 102.00	100.09
Sample	ID			
Tested	02/04/25			
Tester	BKIR			
Recorde	d 02/04/25			
QCSP47:STANDARD	IZATION	YES		YES
VARIOUS CYANIDE		QCSP47		
Sample	ID			
Tested	02/04/25			
Tester	BKIR			
Recorde	d 02/04/25			
EXPIRATION DATE		MM/DD/YY		03/09/25
		MM/DD/YYYY		
Sample				
Tested	02/04/25			
Tester	BKIR			
Recorde				
SHELF LIFE OF PRODUCT		SHELF LIFE		1 MONTH
		TIME		
Sample				
Tested	02/04/25			
Tester	BKIR			
Recorde				
THIS COA HAS BEEN	N REVISED	YES YES		YES
Sample	ID			
Tested	02/04/25			
Tester	BKIR			
Recorde				
Actual ppm Cyanide	-	100.00 PPM	95.00 - 105.00	100.09
Sample	ID			
Tested	02/04/25			
Tester	BKIR			
Recorde				

Page: 2

Date: 02/04/25 at 11:32 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Lot # 5012738 **Made** 02/04/25

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager