



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500615
HA-21055-69

Certificate of Analysis

PRODUCT: DPD Free Chlorine Reagent

PRODUCT NUMBER: 2105569

LOT NUMBER: A4331

MANUFACTURE DATE: 12/12/2024

DATE OF ANALYSIS: 12/12/2024

TEST	SPECIFICATIONS	RESULTS
pH of the reagent in solution	6.2 to 6.5	6.41
Percent recovery for a 2.5 ppm standard. Chlorine concentration determined using DPD compared to the actual concentration.	93 to 107 %	95.6 %
Hardness Reagent Blank: Absorbance of a 1000 ppm as CaCO ₃ hardness standard vs DI water measured at 530 nm in 1 cm cells.	0 to 0.009 abs	0.0010 abs

The expiration date is Dec 2029

Certified by: *Scott Ala*

Analytical Services Chemist