

500611  
HA-1995-32



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

### *Certificate of Analysis*

**PRODUCT:** Molybdate 3 Reagent for Silica

**PRODUCT NUMBER:** 199526

**LOT NUMBER:** A4305

**MANUFACTURE DATE:** 11/19/2024

**DATE OF ANALYSIS:** 11/19/2024

TEST	SPECIFICATIONS	RESULTS
Concentration of a treated 1.0 ppm Silica standard measured at 815 nm in a one inch cell.	0.92 to 1.08 abs	0.970 abs
Absorbance of a treated deionized water sample measured at 815 nm in a one inch cell vs. untreated water.	0 to 0.007 abs	0.0060 abs

The expiration date is Nov 2029

Certified by: *Scott Abo*

Analytical Services Chemist