

500631
HA-26958-53



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

Certificate of Analysis

PRODUCT: APA Reagent 1 Hardness Indicator Solution

PRODUCT NUMBER: 2695853

LOT NUMBER: A4256

MANUFACTURE DATE: 09/23/2024

DATE OF ANALYSIS: 09/24/2024

TEST	SPECIFICATIONS	RESULTS
pH of the solution	5.5 to 6.5	5.90
Difference in absorbances measured in one cm cells at 520 nm between a reagent blank containing EDTA and a reagent blank.	0 to 0.06 abs	0.005 abs

The expiration date is Sep 2025

Certified by: *Scott Als*

Analytical Services Chemist