



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500604
HA-14070-99

Certificate of Analysis

PRODUCT: DPD Free Chlorine Reagent

PRODUCT NUMBER: 1407099

LOT NUMBER: A4331

MANUFACTURE DATE: 12/13/2024

DATE OF ANALYSIS: 12/16/2024

TEST	SPECIFICATIONS	RESULTS
Percent recovery for a 5.0 ppm standard. Chlorine concentration determined using DPD compared to the actual concentration.	93 to 107 %	103.4 %
Hardness Blank: 1000 ppm as Calcium Carbonate Hardness standard vs DI water measured at 530 nm in 1 cm cells.	0 to 0.009 abs	0.0020 abs
Percent recovery for a 2.5 ppm standard. Chlorine concentration determined using DPD compared to the actual concentration.	93 to 107 %	96.5 %
pH of the reagent in 50 mL of DI water.	6.2 to 6.5	6.37

The expiration date is Dec 2029

Certified by: *Scott Als*

Analytical Services Chemist