



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500633  
HA-67748-02

### Certificate of Analysis

**PRODUCT:** 5500sc Reagent 1 Silica

**PRODUCT NUMBER:** 6774802

**LOT NUMBER:** A4326

**MANUFACTURE DATE:** 12/10/2024

**DATE OF ANALYSIS:** 12/06/2024

| TEST                                                                                      | SPECIFICATIONS   | RESULTS   |
|-------------------------------------------------------------------------------------------|------------------|-----------|
| Concentration of a treated 1.0 ppm Silica standard measured at 815 nm in a one inch cell. | 0.92 to 1.08 ppm | 0.990 ppm |
| Series 5500 Blank (ppb)                                                                   | 0 to 5 ppb       | 2.4 ppb   |
| 31201 Instrument Blank (ppb)                                                              | 0 to 5 ppb       | 1.9 ppb   |
| Lab Blank (ppb)                                                                           | 0 to 6 ppb       | 2.2 ppb   |

The expiration date is Nov 2026

Certified by: *Scott Als*

Analytical Services Chemist