



Be Right™

Loveland, CO 80539

(970) 669-3050

500906
HA-21412-99

An ISO 9001 Certified Company

Certificate of Analysis

PRODUCT: Triazole Reagent

PRODUCT NUMBER: 2141299

LOT NUMBER: A4237

MANUFACTURE DATE: 09/11/2024

DATE OF ANALYSIS: 09/12/2024

TEST	SPECIFICATIONS	RESULTS
pH of the powder in solution	5 to 5.2	5.00
Function Test: Absorbance of a 10 ppm benzotriazole standard measured at 425 nm in inch cells vs deionized water.	0.643 to 0.725 abs	0.7200 abs
Reagent Blank: Absorbance measured at 425 nm in inch cells vs deionized water.	0 to 0.013 abs	0.0010 abs

The expiration date is Aug 2027

Certified by: *Scott Als*

Analytical Services Chemist