



Be Right™

An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

500606
HA-1816-32

Certificate of Analysis

PRODUCT: Sulfide 1 Reagent

PRODUCT NUMBER: 181632

LOT NUMBER: A4318

MANUFACTURE DATE: 11/21/2024

DATE OF ANALYSIS: 11/27/2024

TEST	SPECIFICATIONS	RESULTS
Function Test: Absorbance of a standard with a sulfide concentration of 0.5 ppm measured in 1 inch cells vs the reagent blank.	0.899 to 1.121 abs	0.9330 abs
Reagent Blank: Absorbance measured at 655 nm in 1 inch cells vs deionized water.	0 to 0.031 abs	0.0140 abs
Function Test: Absorbance of a standard with a sulfide concentration of 1.0 ppm measured in 1 inch cells vs the reagent blank.	1.828 to 2.214 abs	1.9210 abs

The expiration date is Jul 2028

Certified by: *Scott Abo*

Analytical Services Chemist